

Anti-EBV-EA-D ELISA (IgG)

Test instruction

















ORDER NO.	ANTIBODIES AGAINST	IG CLASS	SUBSTRATE	FORMAT
EI 2795-9601 G	Epstein-Barr virus early antigen diffuse (EBV-EA-D)	IgG	Ag-coated microplate wells	96 x 01 (96)

Indication: The ELISA test kit provides semiquantitative or quantitative in vitro determination of human antibodies of the immunoglobulin class IgG against Epstein-Barr virus early antigen diffuse (EBV-EA-D) in serum or plasma to support the diagnosis of infectious mononucleosis.

Application: Since early EBV antigens (early antigen, EA) are expressed in the lytic replication phase, the detection of specific antibodies using the EUROIMMUN Anti-EBV-EA-D ELISA (IgG) (alternatively IgM) can contribute to supporting the diagnosis of active EBV infections. Results should always be interpreted within the context of clinical symptoms and with respect to further laboratory diagnostic analyses.

Principles of the test: The test kit contains microtiter strips each with 8 break-off reagent wells coated with EBV-EA-D. In the first reaction step, diluted patient samples are incubated in the wells. In the case of positive samples, specific IgG antibodies (also IgA and IgM) will bind to the antigens. To detect the bound antibodies, a second incubation is carried out using an enzyme-labelled anti-human IgG (enzyme conjugate), which is capable of promoting a colour reaction.

Contents of the test kit:

Description	Colour	Format	Symbol
1. Microplate wells coated with antigens 12 microplate strips each containing 8 individual break-off wells in a frame, ready for use	---	12 x 8	
2. Calibrator 1 200 RU/ml (IgG, human), ready for use	dark red	1 x 2.0 ml	
3. Calibrator 2 20 RU/ml (IgG, human), ready for use	red	1 x 2.0 ml	
4. Calibrator 3 2 RU/ml (IgG, human), ready for use	light red	1 x 2.0 ml	
5. Positive control (IgG, human), ready for use	blue	1 x 2.0 ml	
6. Negative control (IgG, human), ready for use	green	1 x 2.0 ml	
7. Enzyme conjugate peroxidase-labelled anti-human IgG (rabbit), ready for use	green	1 x 12 ml	
8. Sample buffer ready for use	light blue	1 x 100 ml	
9. Wash buffer 10x concentrate	colourless	1 x 100 ml	
10. Chromogen/substrate solution TMB/H ₂ O ₂ , ready for use	colourless	1 x 12 ml	
11. Stop solution 0.5 M sulphuric acid, ready for use	colourless	1 x 12 ml	
12. Test instruction	---	1 booklet	
13. Quality control certificate	---	1 protocol	
 Lot description			 Storage temperature
 In vitro diagnostics medical device			 Unopened usable until



Preparation and stability of the reagents

Note: All reagents must be brought to room temperature (+18°C to +25°C) approx. 30 minutes before use. After first use, the reagents are stable until the indicated expiry date if stored at +2°C to +8°C and protected from contamination, unless stated otherwise below.

- **Coated wells:** Ready for use. Tear open the resealable protective wrapping of the microplate at the recesses above the grip seam. Do not open until the microplate has reached room temperature to prevent the individual strips from moistening. Immediately replace the remaining wells of a partly used microplate in the protective wrapping and tightly seal with the integrated grip seam (Do not remove the desiccant bag).
Once the protective wrapping has been opened for the first time, the wells coated with antigens can be stored in a dry place and at a temperature between +2°C and +8°C for 4 months.
- **Calibrators and controls:** Ready for use. The reagents must be mixed thoroughly before use.
- **Enzyme conjugate:** Ready for use. The enzyme conjugate must be mixed thoroughly before use.
- **Sample buffer:** Ready for use.
- **Wash buffer:** The wash buffer is a 10x concentrate. If crystallisation occurs in the concentrated buffer, warm it to +37°C and mix well before diluting. The quantity required should be removed from the bottle using a clean pipette and diluted with deionised or distilled water (1 part reagent plus 9 parts distilled water).
For example: For 1 microplate strip, 5 ml concentrate plus 45 ml water.
The ready-for-use diluted wash buffer is stable for 4 weeks when stored at +2°C to +8°C and handled properly.
- **Chromogen/substrate solution:** Ready for use. Close the bottle immediately after use, as the contents are sensitive to light. ☼ The chromogen/substrate solution must be clear on use. Do not use the solution if it is blue coloured.
- **Stop solution:** Ready for use.

Storage and stability: The test kit has to be stored at a temperature between +2°C and +8°C. Do not freeze. Unopened, all test kit components are stable until the indicated expiry date.

Waste disposal: Patient samples, calibrators, controls and incubated microplate strips should be handled as infectious waste. All reagents must be disposed of in accordance with local disposal regulations.

Warning: The calibrators and controls of human origin have tested negative for HBsAg, anti-HCV, anti-HIV-1 and anti-HIV-2. Nonetheless, all materials should be treated as being a potential infection hazard and should be handled with care. Some of the reagents contain sodium azide in a non-declarable concentration. Avoid skin contact.

Preparation and stability of the patient samples

Samples: Human serum or EDTA, heparin or citrate plasma.

Stability: Patient samples to be investigated can generally be stored at +2°C to +8°C for up to 14 days. Diluted samples should be incubated within one working day.

Sample dilution: Patient samples are diluted **1:101** in sample buffer. For example: dilute 10 µl of sample in 1.0 ml sample buffer and mix well by vortexing (sample pipettes are not suitable for mixing).

NOTE: Calibrators and controls are prediluted and ready for use, do not dilute them.



Incubation

For **semiquantitative analysis** incubate **calibrator 2** along with the positive and negative controls and patient samples. For **quantitative analysis** incubate **calibrators 1, 2 and 3** along with the positive and negative controls and patient samples.

(Partly) manual test performance

Sample incubation: (1st step)

Transfer 100 µl of the calibrators, positive and negative controls or diluted patient samples into the individual microplate wells according to the pipetting protocol. Incubate for **30 minutes** at room temperature (+18°C to +25°C).

Washing:

Manual: Empty the wells and subsequently wash 3 times using 300 µl of working strength wash buffer for each wash.

Automatic: Wash the reagent wells 3 times with 450 µl working strength wash buffer (program setting: e.g. TECAN Columbus Washer "Overflow Mode").

Leave the wash buffer in each well for 30 to 60 seconds per washing cycle, then empty the wells. After washing (manual and automated tests), thoroughly dispose of all liquid from the microplate by tapping it on absorbent paper with the openings facing downwards to remove all residual wash buffer.

Note: Residual liquid (> 10 µl) remaining in the reagent wells after washing can interfere with the substrate and lead to false low extinction readings. Insufficient washing (e.g., less than 3 wash cycles, too small wash buffer volumes, or too short residence times) can lead to false high extinction readings.

Free positions on the microplate strip should be filled with blank wells of the same plate format as that of the parameter to be investigated.

Conjugate incubation: (2nd step)

Pipette 100 µl of enzyme conjugate (peroxidase-labelled anti-human IgG) into each of the microplate wells. Incubate for **30 minutes** at room temperature (+18°C to +25°C).

Washing:

Empty the wells. Wash as described above.

Substrate incubation: (3rd step)

Pipette 100 µl of chromogen/substrate solution into each of the microplate wells. Incubate for **15 minutes** at room temperature (+18°C to +25°C) (protect from direct sunlight).

Stopping:

Pipette 100 µl of stop solution into each of the microplate wells in the same order and at the same speed as the chromogen/substrate solution was introduced.

Measurement:

Photometric measurement of the colour intensity should be made at a wavelength of 450 nm and a reference wavelength between 620 nm and 650 nm **within 30 minutes of adding the stop solution**. Prior to measuring, slightly shake the microplate to ensure a homogeneous distribution of the solution.



Test performance using fully automated analysis devices

Sample dilution and test performance are carried out fully automatically using an analysis device. The incubation conditions programmed in the respective software authorised by EUROIMMUN may deviate slightly from the specifications given in the ELISA test instruction. However, these conditions were validated in respect of the combination of the EUROIMMUN Analyzer I, Analyzer I-2P or the DSX from Dynex and this EUROIMMUN ELISA. Validation documents are available on enquiry.

Automated test performance using other fully automated, open-system analysis devices is possible. However, the combination should be validated by the user.

Pipetting protocol

	1	2	3	4	5	6	7	8	9	10	11	12
A	C 2	P 6	P 14	P 22			C 1	P 4	P 12	P 20		
B	pos.	P 7	P 15	P 23			C 2	P 5	P 13	P 21		
C	neg.	P 8	P 16	P 24			C 3	P 6	P 14	P 22		
D	P 1	P 9	P 17				pos.	P 7	P 15	P 23		
E	P 2	P 10	P 18				neg.	P 8	P 16	P 24		
F	P 3	P 11	P 19				P 1	P 9	P 17			
G	P 4	P 12	P 20				P 2	P 10	P 18			
H	P 5	P 13	P 21				P 3	P 11	P 19			

The pipetting protocol for microtiter strips 1 to 4 is an example for the **semiquantitative analysis** of 24 patient samples (P 1 to P 24).

The pipetting protocol for microtiter strips 7 to 10 is an example for the **quantitative analysis** of 24 patient samples (P 1 to P 24).

The calibrators (C 1 to C 3), the positive (pos.) and negative (neg.) controls, and the patient samples have each been incubated in one well. The reliability of the ELISA test can be improved by duplicate determinations for each sample.

The wells can be broken off individually from the strips. This makes it possible to adjust the number of test substrates used to the number of samples to be examined and minimises reagent wastage.

Both positive and negative controls serve as internal controls for the reliability of the test procedure. They should be assayed with each test run.

Calculation of results

Semiquantitative: Results can be evaluated semiquantitatively by calculating a ratio of the extinction of the control or patient sample over the extinction of the calibrator 2. Calculate the ratio according to the following formula:


$$\frac{\text{Extinction of the control or patient sample}}{\text{Extinction of calibrator 2}} = \text{Ratio}$$

EUROIMMUN recommends interpreting results as follows:

Ratio <0.8:	negative
Ratio ≥0.8 to <1.1:	borderline
Ratio ≥1.1:	positive



Quantitative: The standard curve from which the concentration of antibodies in the patient samples can be taken is obtained by point-to-point plotting of the extinction readings measured for the 3 calibration sera against the corresponding units (linear/linear). Use "point-to-point" plotting for calculation of the standard curve by computer. The following plot is an example of a typical calibration curve. Please do not use this curve for the determination of antibody concentrations in patient samples.



If the extinction for a patient sample lies above the extinction of calibrator 1 (corresponding to 200 RU/ml), the result should be reported as ">200 RU/ml". It is recommended that the sample be re-tested at a dilution of e.g. 1:400. The result in RU/ml read from the calibration curve for this sample must then be multiplied by factor 4.

The upper limit of the normal range of non-infected persons (**cut-off value**) recommended by EUROIMMUN is **20 relative units (RU)/ml**. EUROIMMUN recommends interpreting results as follows:

<16 RU/ml:	negative
≥16 to <22 RU/ml:	borderline
≥22 RU/ml:	positive

For duplicate determinations the mean of the two values should be taken. If the two values deviate substantially from one another, EUROIMMUN recommends retesting the samples.

A negative serological result does not exclude an infection. Particularly in the early phase of an infection, antibodies may not yet be present or are only present in such small quantities that they are not detectable. In case of a borderline result, a secure evaluation is not possible. If there is a clinical suspicion and a negative test result, we recommend clarification by means of other diagnostic methods and/or the serological investigation of a follow-up sample. A positive result indicates that there has been contact with the pathogen. In the determination of pathogen-specific IgM antibodies, polyclonal stimulation of the immune system or antibody persistence may affect the diagnostic relevance of positive findings. Significant IgG titer increases (exceeding factor 2) and/or seroconversion in a follow-up sample taken after at least 7 to 10 days can indicate an acute infection. To investigate titer changes, sample and follow-up sample should be incubated in adjacent wells of the ELISA microplate within the same test run. For diagnosis, the clinical picture of the patient always needs to be taken into account along with the serological findings.

Test characteristics

Calibration: As no international reference serum exists for antibodies against EBV-EA-D, the calibration is performed in relative units (RU).

For every group of tests performed, the extinction readings of the calibrators and the relative units and/or ratios determined for the positive and negative controls must lie within the limits stated for the relevant test kit lot. A quality control certificate containing these reference values is included. If the values specified for the controls are not achieved, the test results may be inaccurate and the test should be repeated.



The binding activity of the antibodies and the activity of the enzyme used is temperature-dependent. It is therefore recommended using a thermostat in all three incubation steps. The higher the room temperature (+18°C to +25°C) during incubation steps, the greater will be the extinction. Corresponding variations apply also to the incubation times. However, the calibrators are subject to the same influences, with the result that such variations will be largely compensated in the calculation of the result.

Antigen: The microplate wells were coated with the recombinant Epstein-Barr virus early antigen diffuse. The protein was expressed in *E. coli* and the molecular weight is 45 kDa.

Linearity: The linearity of the Anti-EBV-EA-D ELISA (IgG) was determined by assaying at least 4 serial dilutions of different patient samples. The Anti-EBV-EA-D ELISA (IgG) is linear at least in the tested concentration range (2 RU/ml to 158 RU/ml).

Detection limit: The lower detection limit is defined as the mean value of an analyte-free sample plus three times the standard deviation and is the smallest detectable antibody titer. The lower detection limit of the Anti-EBV-EA-D ELISA (IgG) is 0.8 RU/ml.

Cross reactivity: This ELISA showed no cross reactivity.

Interference: Haemolytic, lipaemic and icteric samples showed no influence on the result up to a concentration of 10 mg/ml for haemoglobin, 20 mg/ml for triglycerides and 0.4 mg/ml for bilirubin in this ELISA.

Reproducibility: The reproducibility of the test was investigated by determining the intra- and inter-assay coefficients of variation (CV) using 3 samples. The intra-assay CVs are based on 20 determinations and the inter-assay CVs on 4 determinations performed in 6 different test runs.

<i>Intra-assay variation, n = 20</i>		
Sample	Mean value (RU/ml)	CV (%)
1	32	4.0
2	102	4.1
3	142	3.1

<i>Inter-assay variation, n = 4 x 6</i>		
Sample	Mean value (RU/ml)	CV (%)
1	33	4.5
2	107	7.0
3	146	5.2

Sensitivity and specificity: 35 clinically pre-characterised patient samples (INSTAND) were investigated with the EUROIMMUN Anti-EBV-EA-D ELISA (IgG). The sensitivity amounted to 100% with a specificity of 100%. Borderline results were not included in the calculation.

n = 35		INSTAND		
		positive	borderline	negative
EUROIMMUN Anti-EBV-EA-D ELISA (IgG)	positive	3	0	0
	borderline	0	0	0
	negative	0	2	30

Reference range: The levels of the anti-EBV-EA-D antibodies (IgG) were analysed with this EUROIMMUN ELISA in a panel of 297 healthy blood donors. With a Cut-off of 20 RU/ml, 5% of the blood donors were anti-EBV-EA-D positive (IgG).



Clinical significance

Epstein-Barr virus (EBV) is one of the most widely distributed human-pathogenic herpes viruses. The virus is transmitted by smear infection, but also by blood transfusions or organ transplants. EBV is the causative agent of infectious mononucleosis (Pfeiffer's disease), a febrile disease usually accompanied by pharyngitis and lymphadenopathy, frequently by hepatosplenomegaly and more rarely by an exanthema. When first infection occurs in childhood the disease often proceeds without symptoms. In industrial countries mainly adolescents or young adults become infected, often leading to manifest disease. EBV infections are also found in connection with the pathogenesis of malignant lymphoma (endemic form of Burkitt's lymphoma in Africa) and nasopharyngeal carcinoma (NPC, especially widespread in South-East Asia). NPC is the third most frequent malignant tumour in southern China. Since 2011 the evidence has been increasing that EBV infection is associated with a high risk of multiple sclerosis (MS) or an aggravation of MS.

The main goal of EBV diagnostics in people with a healthy immune system is to differentiate between an acute and a past infection. Various serological methods are used for this. The immune system of healthy persons can quickly suppress a reactivation of the virus. However, in immunosuppressed patients (e.g. those on immunosuppressive therapy after organ transplantation or with an HIV infection) EBV can spread uncontrollably and cause severe lymphoproliferative diseases. In such cases it is diagnostically very important to also determine the viral load, for which PCR (polymerase chain reaction) is normally used.

Infectious mononucleosis must be differentiated from cytomegalic inclusion body disease and toxoplasmosis and, in the case of atypical progress, also from HIV or other infections.

In pregnancy, EBV can cause infection of the placenta, leading to damage to the foetal heart, eyes and liver. In children, accompanying infections of the kidney have been observed with symptoms from microscopic haematuria to acute kidney failure.

The immune response to an EBV infection is characterised by the successive formation of antibodies against EBV capsid, EBV nuclear and EBV early antigens.

In the early phase of the disease IgM and IgG antibodies against the viral capsid antigen (CA) can be detected. A positive Anti-EBV-CA (IgM) result is the classical marker of acute infection. IgG antibodies against early antigen are produced later in the acute phase and decrease to a non-detectable level after three to six months. Anti-CA IgG antibody levels persist lifelong. Around six to eight weeks after infection, antibodies against EBNA are produced. The presence of anti-EBNA antibodies indicates a past infection.

Literature references

1. Ascherio A, Munger KL. **99th Dahlem conference on infection, inflammation and chronic inflammatory disorders: Epstein-Barr virus and multiple sclerosis: epidemiological evidence.** Clin Exp Immunol 160 (2010) 120-124.
2. Avgil M, Ornoy A. **Herpes simplex virus and Epstein-Barr virus infections in pregnancy: consequences of neonatal or intrauterine infection.** Reprod Toxicol 21 (2006) 436-445.
3. Bauer G. **Simplicity through complexity: immunoblot with recombinant antigens as the new gold standard in Epstein-Barr virus serology.** Clin Lab 47 (2001) 223-230.
4. Bertrand KA, Birmann BM, Chang ET, Spiegelman D, Aster JC, Zhang SM, Laden F. **A prospective study of Epstein-Barr virus antibodies and risk of non-Hodgkin lymphoma.** Blood 116 (2010) 3547-3553.
5. De Paschale M, Clerici P. **Serological diagnosis of Epstein-Barr virus infection: Problems and solutions.** World J Virol 1 (2012) 31-43.
6. Gärtner B, Hess R, Brandt D, Kruse A, Rethwilm A, Roemer K, Mueller-Lantzsch N. **Evaluation of four commercially available Epstein-Barr virus enzyme immunoassays with an immunofluorescence assay as the reference method.** Clin Diagn Lab Immunol 10 (2003) 78-82.

7. Goldstein BL, Chibnik LB, Karlson EW, Costenbader KH. **Epstein-Barr virus serologic abnormalities and risk of rheumatoid arthritis among women.** Autoimmunity 45 (2012) 161-168.
8. Levin LI, Chang ET, Ambinder RF, Lennette ET, Rubertone MV, Mann RB, Borowitz M, Weir EG, Abbondanzo SL, Mueller NE. **Atypical prediagnosis Epstein-Barr virus serology restricted to EBV-positive Hodgkin lymphoma.** Blood 120 (2012) 3750-3755.
9. EUROIMMUN AG. Lipkowski M, Viertel V, Steller U, Fechner K, Koriath S, Stöcker W, Rohwäder E. **Monospecific substrates coated with gp125 and p19 antigens can improve the serologic diagnosis of EBV infections by IIFT.** International Journal of Medical Microbiology 298S2 (Suppl. 45) 10 (2008).
10. Maeda E, Akahane M, Kiryu S, Kato N, Yoshikawa T, Hayashi N, Aoki S, Minami M, Uozaki H, Fukayama M, Ohtomo K. **Spectrum of Epstein-Barr virus-related diseases: a pictorial review.** Jpn J Radiol 27 (2009) 4-19.
11. Mueller NE, Lennette ET, Dupnik K, Birmann BM. **Antibody titers against EBNA1 and EBNA2 in relation to Hodgkin lymphoma and history of infectious mononucleosis.** Int J Cancer 130 (2012) 2886-2891.
12. Pakpoor J, Giovannoni G, Ramagopalan SV. **Epstein-Barr virus and multiple sclerosis: association or causation?** Expert Rev Neurother 13 (2013) 287-297.
13. Reiber H, Lange P. **Virus-spezifische Antikörper in Liquor und Serum.** ELISA-Analytik und Auswertung mittels Antikörper-Index und Quotientendiagramm. Lab Med 15 (1991) 204-207.
14. EUROIMMUN AG. Sonnenberg K, Gaertner B, Steinhagen K, Meyer W, Ossendorf C, Rohwäder E, Scheper T, Müller-Kunert E, Schlumberger W, Stöcker W. **Reliable EBV diagnostic using IIF, ELISA and western blot.** Poster zum Europäischen Virologie-Kongress, Glasgow 2000. J Clin Virol 18 (2000) 92.
15. Sonnenberg* K, Hinrichs E, Müller-Kunert* E, Schlumberger* W, Stöcker* W. (*EUROIMMUN AG). **Einsatz der BIOCHIP-Technologie in der Serologie von Influenza-, EBV- und Legionella-Infektionen.** Poster zum 3. Deutschen Kongress für Infektions- und Tropenmedizin, 15. bis 18. März 1995 in Berlin-Hohenschönhausen. Abstrakt im Kongressband (1995).
16. Stöcker* W, Geusendam G, Schmitz H, Mohr H, Berndt H. (*EUROIMMUN AG). **Rational microanalysis technic in the detection of antibodies to HTLV-III, CMV and EBV with indirect immunofluorescence.** Beitr Infusionsther Klin Ernähr 15 (1986) 51–57.
17. Tang JW, Rohwäder* E, Chu IMT, Tsang RKY, Steinhagen* K, Yeung ACM, To KF, Chan PKS. (*EUROIMMUN AG). **Evaluation of Epstein-Barr virus antigen-based immunoassays for serological diagnosis of nasopharyngeal carcinoma.** J Clin Virol 40 (2007) 284-288.